



# Certificate of Analysis

---

1.88005.1000 CombiTitrant 5 one-component reagent for volumetric Karl Fischer titration 1  
ml  $\triangleq$  ca. 5 mg H<sub>2</sub>O Aquastar®  
Batch HX32322605

---

	Spec. Values		Batch Values	
Efficiency	$\geq 5$	mg/ml	5.33	mg/ml

The concentration (efficiency) of Karl Fischer titrants decreases over time. Therefore the titer, being a corrective factor that compensates for this effect as well as variations of equipment, personal handling and laboratory conditions, must be determined directly after opening a new bottle and should be verified at regular intervals.

Date of release (DD.MM.YYYY) 01.06.2023  
Minimum shelf life (DD.MM.YYYY) 30.06.2026

Ayfer Yildirim  
Responsible laboratory manager quality control

---

This document has been produced electronically and is valid without a signature.